

WATER BATH SC-20

User Manual



Please read operating manual before installation and operation.

Drawell International Technology Limited

Chongqing Drawell Instrument Co., Ltd.

Shanghai Drawell Scientific Instrument Co.,Ltd.

Address : Suite 2705, Building No.12, Shiyou Road No.1, Yuzhong District,
Chongqing China.

Homepage : www.drawell.com.cn

Tel : 0086-023-63268643

Email : sales05@drawell.com.cn

Characteristics:

1. Low Temperature Coolant circulator is used wind-cool-type and complete-close-compressor set to refrigerate. By microprocessor control, cooling systems automatically start, time-lapse turn off, with over-heat and over-current protection
2. The Circulator may export the coolant from reservoir, cool or maintain the temperature of out side lab vessel or establish second constant temperature field.
3. Microprocessor controls the time switch from 0-100 hours.
4. Dual windows, Red and Green LED display the set temperature and the actual temperature, respectively. The readout accuracy is $\pm 0.1^{\circ}\text{C}$, and the stability is $\pm 0.1^{\circ}\text{C}$.
5. Microprocessor-controlled-lock function may lock all parameters.
6. Having Easy-to-use microprocessor control touch keys.

Technical Parameters:

Model	Temp. Range ($^{\circ}\text{C}$)	V/AC	Capacity (L)	Accuracy ($^{\circ}\text{C}$)	Pump flow (L/min)	Drain	Inner bath size	Bath open size
SC-20	Room+8-100	220V/50HZ	20	± 0.1	0-15	yes	500*300*150mm	310*280mm

Operation Process:


Add liquid media in bath, with liquid level no less than 30mm lower than work bench.

1. Selection of liquid media
 - A) When work temperature is below 5°C , liquid media shall be alcohol.
 - B) When work temperature is between 5°C - 80°C , normally liquid media shall be pure water.
 - C) When work temperature is between 80°C - 90°C , normally liquid media shall be 15% glycerin water solvent.
 - D) When work temperature over 90°C , normally liquid media shall be Silicone oil
2. Connection of cycling pump:
 - A) For connection of internal and external cycling pumps, the user only need to connect liquid outlet tube to liquid inlet tube with hose which will be provided with the machine).
 - B) Connect external cycle of external cycling pump, connect liquid outlet tube to inlet of vessel outside the bath, and connect liquid inlet tube to outlet vessel outside the bath.
(Note: tube on the left side of front panel of instrument is liquid inlet tube, and the one on the rack panel of the instrument is liquid outlet tube)


3. Put on the power (on rack control panel)

4. Instrument shall be operated as follows:

Circulating details



Picture Inner Loop 1.1




Picture Outer Loop 1.2

- Inner loop: connect with outlet and inner circulating via hose
- Outer loop: The outlet and inner circulating connected to the outer container respectively, set up the second temperature container (Picture 1.2)
- Drain: When finish all process or need change water, open the drain valve, water can achieved, under the water must a water container, or use the water pipe connected to the gutter



Display (As picture 2) :

Interface:









Picture 2

Function key :

1.  Refrigeration light: When the light flashing refrigeration is on, the cooling function start working after 2-3 minutes, Refrigeration lights long time bright
2.  Circulating pump light: when long press the RUN key button machine, circulating

pump auto working, when shut off, the light dark


3.  Heating light, start heat function, the light on, or it's off
4.  Water level light: the light tip groove liquid lever cannot under working table 30mm
5.  RUN key : Long press 5 second, the the device start or stop working
6.  SET key : TEM Setting and TEM Measure set confirm, and amend etc.
7.  LEFT MOVE KEY: SETTING data move key
8.  UP and DOWN KEY : SETTING DATA "+ "or" -"Long press UP key ,can be check environment temperature

Notice : if TEM Measure area display “0000,” indicate sensor input or the signal over measurement limit ,if display “ 0000 ”, indicate sensor short circuit or the input signal is lower than the lower limited range when the input signal is beyond the scope of range ,buzzer ,press any key be muted.

Data setting

- A) Please join the bath in liquid medium (such as alcohol).
- B) Turn on the power

1、When turn on the device , (as picture 3)display CP, the first row it's sing point control ,the second empty ,the third it's number 1 ,means reach setting value ,if it's 0 ,that means electric start timing)buzzer alarm ,automatically after 3 ring back to standby model as picture 3




Picture 3

2、Setting temperature and time:

1) Press SET key , , TEM Setting area temperature value bottom flash, via pressing(LEFT) shift, increased (UP), and reduce key (DOWN), to SET the required temperature.

2) Press SET key enter into Timing area value bottom flash, via pressing(LEFT) shift, increased (UP), and reduce key (DOWN), to SET the required time, press SET key ,then save and return to standby model (see picture 4).




Picture 4 Temperature and time setting

3、Timing function

When the Timing is set to 0, instrument cancel the Timing function and instrument always running; When Timing set not 0, the meter is timing function, when the meter running time, Timing show END, buzzer, instrument to stop working, and press any key can be muted.

The device work start/stop:

Long press the RUN more than 3 seconds, compressor buzzing, cooling and circulating pump began to work. If you want to stop working and long press the RUN button will stop working, the buzzer alarm, press any key mute.(see picture 5) :



Picture 5 Timing function

Tech Tip

1. Before using, should add full liquid medium.
2. The power supply should be match with the technical parameter. The power should

be greater then or equal to the total power of the machine. Ground wire should be used
(Connect the snail located at the left bottom on the rear side) .

3. The machine should be put in dry and drafty place, rear and two sides 400mm apart
from other articles.

4. After done, all switches should be in turn off Pull out plug, dry and clean bath

Packing list :

1. Main machine	1set
2. Power supply wire	1pc
3. Rubber tube	1pc
4. Cover	1pc
5. Operational Manual	1copy

Drawell International Technology Limited

**Shanghai Drawell Scientific Instrument
Co.,Ltd.**

Chongqing Drawell Instrument Co.,Ltd.



Add: Suite 2705, Building No.12, Shiyou Road No.1, Yuzhong District,
Chongqing, China

Tel: 0086-023-63268643

Web : www.drawell.com.cn

Email : sales05@drawell.com.cn